

Technical Data Sheet

Edition Date: 2013-02-28 **Revision Date:** 2018-06-22

Revised by: Annette Roy, QA Coordinator

Product Name: Amplitube[™] Catalogue No.: T320-1 Series

This document replaces any previous version

1. Product Description:

> PCR Reaction Strip of 8 Tubes 0.2ml: Non-sterile, disposable tubes only

> Compatible with

- T321-1 Series Strip of 8 Dome Caps
- T321-2 Series Strip of 8 Flat Caps
- > Available colors:
 - · Natural, Blue, Green, Red and Yellow

2. Packaging:

Case: 125 units in a package (tamperproof resealable bag)

3. Product Specifications:

- Material:
 - Polypropylene
- Certified RNase, DNase, Pyrogen and DNA Free
- Strip of 8 tubes of 0.2ml, thin wall construction
- Compatible with multi-channel pipettor
- > Temperature range: -80°C to +121°C
- Autoclavable at +121°C for up to 30 minutes

4. Standards and Conformity:

➤ **ISO 2859-1:** Sampling and inspection procedures

FDA: Resin conforms to FDA 21 CFR 177.1520

CONEG / RoHS:
Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

➤ BSE / TSE: Material is BSE / TSE Free

➤ LATEX: Material is Latex Free

Document Template: FRM-106 Rev00 Technical Data Sheet Issue Date: 2016-04-26



Technical Data Sheet

Edition Date: 2013-02-28 **Revision Date:** 2018-06-22

Revised by: Annette Roy, QA Coordinator

5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Closure verification
- Evaporation testing in PCR Cycler

6. Traceability:

Lot No. Composition: 8 to 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging, which could compromise the product integrity

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26